



Test attachment for torque screwdrivers

7791-2

Product no. **52110291**
GTIN **4018754178360**
Model **7791-2**


Label.

Test attachment for torque screwdrivers N·m

Properties.

- can be attached to the manual calibration system No. 7791
- the torque screwdriver to be calibrated is inserted into the transducer and fixed in place using the universal centre clamps
- the force can be applied in a controlled manner using the handwheel

Technical drawing.



Technical attributes.

variable height	442-593 mm
Range N·m	up to 10 N·m
Width mm (b)	250 mm
t	351 mm

Logistics data.

Depth mm (IFS)	440
Width mm (IFS)	370
Height mm (IFS)	245
WEEE/ElektroG	nicht ear-pflichtig
Length (packaged, mm)	600

Width (packaged, mm)	400
Height (packaging, mm)	270
Volume (packaged, dm3)	64.8 dm3
Product no.	52110291
Weight (gross, kg)	5,400
Weight PAP (kg)	0,250
Weight PVC (kg)	0,000
GTIN	4018754178360
Country of origin AWR	GERMANY
Region of origin	Nordrhein-Westfalen
Customs tariff no.	90319000
Packing standard	1
Weight (g)	3739 g

Variants.

Product no.	Model No. (ERP)	Description	GTIN
52110291	7791-2	Test attachment for torque screwdrivers N·m	4018754178360

GTIN.



STAHLWILLE Eduard Wille GmbH

Lindenallee 27 · 42349 Wuppertal · Germany · Phone: +49 202 4791-0

info@stahlwille.de · www.stahlwille.com

© STAHLWILLE Eduard Wille GmbH, Wuppertal